



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/698,253	10/31/2003	Manuel A. Villafana	2003-2052.CIP	2949
22476	7590	06/20/2005		
HAUGEN LAW FIRM SUITE 1130 - TCF TOWER 121 SOUTH EIGHTH STREET MINNEAPOLIS, MN 55402			EXAMINER BIANCO, PATRICIA	
			ART UNIT 3762	PAPER NUMBER

DATE MAILED: 06/20/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/698,253

Applicant(s)

VILLAFANNA ET AL.

Examiner

Patricia M. Bianco

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 03 November 2003.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-19 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-19 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☒ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 03 November 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date 05/04; 12/03.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☒ Other: Detailed Action.

DETAILED ACTION

Specification

The disclosure is objected to because of the following informalities: It is the Examiner's position that Applicant has evoked sixth paragraph, means-plus-function language to define Applicant's invention. Therefore the Examiner requires the Applicant to amend the specification pursuant to 37 CFR 1.75(d) and MPEP 608.01(o) to explicitly state, with reference to the terms and phrases of the claim element, what structure, materials, and acts perform the function recited in the claim element. Please note that the MPEP clearly states, "Even if the disclosure implicitly sets forth the structure, materials, or acts corresponding to the means-(or step-) plus-function claim element in compliance with 35 U.S.C. 112, first and second paragraphs, the PTO may still require the applicant to amend the specification pursuant to 37 CFR 1.75(d) and MPEP 608.01(o)...". (Also see **MPEP 2181** (Rev. 1, Feb.2000)) *Wolfensperger*, 302 F.2d at 955, 133 USPQ at 542. Appropriate correction is required. Applicant includes the claim limitations "***coupling means***" and "***means for directing bypass***" in claim 1 and "***flow restricting means***" in claims 8 & 9. Explicit structure is required in the specification for these means-plus-function recitation.

Appropriate correction is required.

Claim Objections

Claims 1, 8, & 9 are objected to because of the following informalities: It is the Examiner's position that Applicant has evoked sixth paragraph, means-plus-function

language to define Applicant's invention. Therefore the Examiner has objected to the claims for the reasons set forth above in the objection to the specification.

Appropriate correction is required.

Claim 3 is objected to because of the following informalities: it appears that a word (perhaps "of") is missing between end and said in line 3 on page 29 of the claim. Appropriate correction is required.

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claim 4 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. The limitation set forth in claim 4 is not further limiting, since it is recited in claim 1, lines 12-14.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 4-9, & 15-19 are rejected under 35 U.S.C. 102(b) as being anticipated by Villafana (6,241,761). Villafana discloses a connector (30) having multiple channels (31/32 50/51 56/57), which can be a delivery or supply conduit. Each has a lumen

Art Unit: 3762

having a circular diameter and therefore communicates with one another through apertures (their respective openings) at their intersection. The multiple channel may be connected in an angular or perpendicular relationship (see figures 2-10). The connector has grooves (38A/B, 40A/B, 41, 36) for connecting to a graft segment. Villafana teaches that the lumen diameter is approximately 2 mm, but can be larger or smaller if appropriate. With respect to the language "configured to be operably connected to" and "for coupling" graft a member to the conduits, Villafana shows that they may be operatively coupled to multiple graft segments (shown, for example, by illustration of the connection to graft (35) in figure 3). Further, they are functional recitations of using the device. The functional recitations have not been given patentable weight because they are in narrative form. In order to be given patentable weight, a functional recitation must be expressed as a "means" for performing the specified function, as set forth in 35 USC § 112, 6th paragraph, and must be supported by recitation in the claim of sufficient structure to warrant the presence of the functional language. *In re Fuller* With respect to claim 9, the recitation has been treated as an intended use recitation. It has been held that a recitation with respect to the manner in which a claimed apparatus is intended to be employed does not differentiate the claimed apparatus from a prior art apparatus satisfying the claimed structural limitations. *Ex parte Masham*, 2 USPQ2d 1647 (1987).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 10-14 are rejected under 35 U.S.C. 103(a) as being unpatentable over Villafana ('761) in view of Davila et al. (2003/0065377). Villafana discloses the invention substantially as claimed, see rejection supra, however, fails to disclose specifically that the a biocompatible coating, such as heparin, may be on at least the inner surfaces of the connector &/or graft. Villafana discloses coating the device if it is made of metal. Davila et al. disclose coated medical devices, such as connectors and grafts, wherein the coating may be heparin. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the connector and graft of Villafana to have a heparin coating, since it has been held to be within the general skill of a

Art Unit: 3762

worker in the art to have heparin coated medical devices when the device is to be implanted in the vasculature. Such a coating prevents thrombosis formation.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1-18 are rejected under the judicially created doctrine of double patenting over claims 1-17, 19, 20, & 23 of Application No. **10/634,200** since the claims, if allowed, would improperly extend the "right to exclude" already granted in the patent.

The subject matter claimed in the instant application is fully disclosed in the patent and is covered by the patent since the patent and the application are claiming common subject matter, a combination of means for directing bypass blood through graft segments.

Furthermore, there is no apparent reason why applicant was prevented from presenting claims corresponding to those of the instant application during prosecution of

the application which matured into a patent. See *In re Schneller*, 397 F.2d 350, 158 USPQ 210 (CCPA 1968). See also MPEP § 804.

Claims 1-12 & 16-18 are rejected under the judicially created doctrine of double patenting over claims 1-14, 16, & 17 of Application No. **10/909,700** since the claims, if allowed, would improperly extend the "right to exclude" already granted in the patent.

The subject matter claimed in the instant application is fully disclosed in the patent and is covered by the patent since the patent and the application are claiming common subject matter, a combination of means for directing bypass blood through graft segments.

Furthermore, there is no apparent reason why applicant was prevented from presenting claims corresponding to those of the instant application during prosecution of the application which matured into a patent. See *In re Schneller*, 397 F.2d 350, 158 USPQ 210 (CCPA 1968). See also MPEP § 804.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Patricia M. Bianco whose telephone number is (571) 272-4940. The examiner can normally be reached on Monday to Friday 9:00-6:30, alternate Fridays off.

Art Unit: 3762

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on (571) 272-4955. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

June 11th, 2005

Patricia M Bianco
Primary Examiner
Art Unit 3762


PATRICIA BIANCO
PRIMARY EXAMINER